


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-MI-P-40

S.N.: 4369

Date: JUNE 8, 1988



- b) Output signal amplitude:
 $\geq +40$ VOLTS
- c) Pulse width:
 ≤ 1.0 NSEC
- d) Rise time:
 NA
- e) Fall time:
 NA
- f) PRF:
 0 TO 1.0 MHz
- g) Jitter, stability:
 OK
- h) Prime power:
 $+135 \pm 5.0$ VOLTS
 5 mA (MAX)

40 db ATTEN
10 VOLTS/DIV
700 KHz PRF