


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-MI-P-40

S.N.: 4371

Date: JUNE 8, 1988



- a) Output signal amplitude:
 $\geq + 40 \text{ VOLTS}$
- b) Pulse width:
 $\leq 1.0 \text{ NSEC}$
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+135 \pm 5.0 VOLTS

5 mA (MAX)

40 db ATTEN

$\therefore 10^0 \text{ VOLTS/DIV}$

700 KHZ PRF