


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-2-PS-P.

S.N.: 4385

Date: JUNE 9, 1988



- a) Output signal amplitude:  
0 TO +350 VOLTS
- b) Pulse width:  
5 TO 100 NSEC
- c) Rise time:  
 $\leq 2$  NSEC
- d) Fall time:  
 $\leq 3$  NSEC
- e) PRF:  
0 TO 5 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V  
50-60 HZ

70 db ATTEN

320 VOLTS/DIV

2 KHZ PRF