


PULSE GENERATOR  
PERFORMANCE CHECK

Model: PM-102-P

S.N.: 4390

Date: JULY 29 88



- a) Output signal amplitude:  
≥ +60 VOLTS
- b) Pulse width:  
≤ 2 NSEC
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:

50 db ATTEN

32 VOLTS/DIV

PRF = 500 KHz

+135 ± 5 VOLTS

5 mA (MAX)

[Signature]