


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVN-1-C-PN-M-0T

S.N.: 4391

Date: JUL 8 1988



- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.2 TO 1.0 NS
- c) Rise time:
 ≤ 100 PSEC
- d) Fall time:
 ≤ 100 PSEC
- e) PRF:
50 TO 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
5 VOLTS/DIV

120/240 V
50-60 Hz