


PULSE GENERATOR
PERFORMANCE CHECK


Model: *ME-V-1411-CPN-0T*

S.N.: *4393*

Date: *JULY 20 1988*



Ⓐ RISE TIME
50 db ATTEN
∴ ≈ 32 VOLTS/DIV
10 KHz



Ⓑ AS Ⓐ BUT
5 μSEL/DIV

- a) Output signal amplitude:
0 TO ±50 VOLTS
- b) Pulse width:
4 TO 100 NSEL
- c) Rise time:
≤ 250 PSEL
- d) Fall time:
≤ 2 NSEL
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

A