


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PLAMP-2-C-M-CA-N*

S.N.: *4396*

Date: *JULY 19 88*



*40 db ATTEN
10 VOLTS/DIV
1.0 MHz*

- a) Output signal amplitude:
0 TO -10 VOLTS
- b) Pulse width:
5.0 TO 100 NSER
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 100 PSEZ
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 HZ*

[Signature]