


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-2.C-P-M*  
S.N.: *4400*  
Date: *July 12 1988*



- a) Output signal amplitude:  
*0 TO +15 VOLTS*
- b) Pulse width:  
*0.2 TO 2.0 nSEC*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 135 pSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*40 db ATTEN*  
*10 VOLTS/DIV*  
*≈ 12 MHz PRF*

*120/240 V*  
*50-60 Hz*