


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-4B-PS-P*

S.N.: *4411*

Date: *July 28 1988*



- b) Output signal amplitude:
0 TO +400 VOLTS
- c) Pulse width:
0.1 TO 100 µS
(EQUALS PULSE PW)
- d) Rise time:
≤ 20 NS
- e) Fall time:
≤ 20 NS
- f) PRF:
MIN PW: 0 TO 1KHz
MAX PW: 0 TO 50Hz
- g) Jitter, stability:
OK
- h) Prime power:
120/240 V
50-60 Hz

R_L = 50 Ω
PRF = 50 Hz

120/240 V
50-60 Hz
A