


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PRR-4B-PS-N*

S.N.: *4412*

Date: *July 18 1988*



- d) Output signal amplitude:
0 TO 400 VOLTS
- e) Pulse width:
0.1 TO 100 nS
- f) Rise time:
(EQUALS DELTA PW)
≤ 20 nS
- g) Fall time:
≤ 20 nS
- h) PRF:
MIN PW: 0 TO 1 kHz
MAX PW: 0 TO 50 kHz
- i) Jitter, stability:
OK
- j) Prime power:

R_L = 50 Ω
PRF = 50 kHz

120/240 V
50-60 kHz

[Signature]