



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVPP-1-PS-P-PN-M-EP-EN-EO*
S.N.: *4414*
Date: *JUNE 23 88*



A out



B out

40 dB ATTEN

10 VOLTS/DIV

1.0 MHz PRF

- a) Output signal amplitude:
0 TO ±10 VOLTS
- b) Pulse width:
0.2 TO 100 NS
- c) Rise time: (TWO RANGES)
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V

50 - 60 Hz