


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AMK-4B-PW-C-P*

S.N.: *4422*

Date: *JUNE 6 1988*



Ⓐ  $\approx 1.0 \text{ kHz}$   
 $R_L = 50 \Omega$



Ⓑ  $\approx 50 \text{ MHz}$   
 $R_L = 50 \Omega$

- a) Output signal amplitude:  
*0 TO +400 V TO*
- b) Pulse width: *50 ns*  
*0.1 TO 100 USEC*
- c) Rise time:  
*≤ 20 USEC*
- d) Fall time:  
*≤ 20 USEC*
- e) PRF:  
*MAX DW: 0 TO 50 MHz*  
*PW ≤ 50 USEC: 0 TO 2.0 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 MHz*

*[Signature]*