


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVI-V-HV3-CPN-0T*

S.N.: *4424*

Date: *JULY 29 88*



Ⓐ *60 db ATTEN*
100 VOLTS/DIV
(RISE TIME)



Ⓑ *AS Ⓐ BUT*
5 nSEC/DIV

- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width:
5 TO 100 nSEC
- c) Rise time:
≤ 200 pSEC
- d) Fall time:
≤ 4 nSEC
- e) PRF:
0 TO 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

- h) OFFSET:
0 TO ± 5 VOLTS
TO 50 Ω

[Signature]