


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM4-1-P*

S.N.: *4426*

Date: *JULY 21 '88*



40 dB ATTEN
10 VOLT/DIV
PRF = 25 MHz

- a) Output signal amplitude: *0 TO +10 VOLTS*
- b) Pulse width: *6200 PSEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:

+24 ± 1.0 VOLT.
500 mA (MAX)

