


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ASMR-2-C-P*

S.N.: *4427*

Date: *JUNE 16 '80*



- a) Output signal amplitude:  
*0 TO +20 VOLTS*
- b) Pulse width:  
*20 TO 200 NSEC*
- c) Rise time: *(20% MAX DUTY CYCLE)*  
 *$\le 3\text{ NSEC}$*
- d) Fall time:  
 *$\le 3\text{ NSEC}$*
- e) PRF: *0 TO 10 MHz*  
*(20% MAX DUTY CYCLE)*
- f) Jitter, stability:
- g) Prime power: *OK*

*40 db ATTEN*  
*10 VOLTS/DIV*  
*1.0 MHz*

*120/240V*  
*50-60 MHz*