



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMMP-3-C-PN-M-EW*
S.N.: *4436*
Date: *AUG 2 1 988*



Ⓐ *40 db ATTEN*
10 VOLTS / DIV
1.0 MHz PRF



Ⓑ *AS Ⓐ BUT*
1 nSEC / DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO ± 20 VOLTS
- b) Pulse width:
8 TO 100 nSEC
- c) Rise time:
≤ 200 pSEC
- d) Fall time:
≤ 300 pSEC
- e) PRF:
0 TO 1.0 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz.