


PULSE GENERATOR
PERFORMANCE CHECK


Model: *MR-B3-e-PA-LIA-N*

S.N.: *4437*

Date: *July 14 '83*



*(A) 60 db ATTEN
100 VOLTS DIV
PRF = 200 KHz*



*(B) $R_L = 50 \Omega$
PRF = 200 KHz*

- a) Output signal amplitude:
0 TO -100 V. (TO 50V)
 - b) Pulse width:
20 TO 100 NSOR
 - c) Rise time:
 ≤ 5 NSOR
 - d) Fall time:
65 NSOR
 - e) PRF:
0 TO 200 KHz
 - f) Jitter, stability:
OK
 - g) Prime power:
*120/240 V
50-60 Hz*
- [Signature]*