


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P*
S.N.: *4438*
Date: *JULY 5 1988*



- a) Output signal amplitude:
25 TO 100 AMP
- b) Pulse width:
≤ 25 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 60 AMPS/PIU
500 Hz.

+15 ± 1.0 VOLT.
200 mA (max).