


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-1-P*

S.N.: *4439*

Date: *JULY 5 1988*



- a) Output signal amplitude:  
*25 TO 100 AMPS*
- b) Pulse width:  
*≤ 25 nSec*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*≈ 60 AMPS/DIV*  
*500 Hz*

*+15 ± 1.0 VOLTS.*  
*200 mA (MAX.)*