


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVB-200*
 S.N.: *4454*
 Date: *JULY 7 1988*



- a) Output signal amplitude: *≥ 20 VOLTS*
- b) Pulse width: *(P-P) 5 nSEC*
- c) Rise time: *(200 MHz)*
- d) Fall time: *NA*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:

→ ← 5 nSEC

∴ T⁻¹ = 200 MHz

50 dB ATTEN

32 VOLTS/DIV

PRF = 500 kHz

+15 ± 1.0 VOLT

200 mA (MAX)

[Signature]