



PULSE GENERATOR
PERFORMANCE CHECK

Model: *MG-200*
 S.N.: *4455*
 Date: *JULY 7 1988*



Ⓐ $\leftarrow \rightarrow$
 5.0 ns
 $\therefore T = 200 \text{ MHz}$



Ⓑ 50 dB ATTEN
 32 VOLTS DIV
 $\text{PRF} = 500 \text{ KHz}$

- a) Output signal amplitude: *~ 20 VOLTS (P-P)*
- b) Pulse width: *5 nSEC*
- c) Rise time: *(200 nite)*
- d) Fall time: *NA*
- e) PRF: *NA*
- f) Jitter, stability: *0 TO 1.0 MHz*
- g) Prime power: *OK*

+15 ± 1.0 VOLT
200 mA (MAX)