


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVO-5A-C-A-CA-CA-T

S.N.: 4457

Date: AUG 16 88



- a) Output signal amplitude:
0 TO +40 AMPS TO
- b) Pulse width: 0.75 μ s
20 TO 150 NSER.
- c) Rise time:
 \leq 5 NSER
- d) Fall time:
 \leq 5 NSER
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

\approx 40 AMPS/DIV

($R_c = 0.75 \mu$
50 dB ATTEN
 \approx 32 VOLTS/DIV)

PRF = 5.0 KHZ

120/240V

50-60 KHZ

[Signature]