


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ARM-1.C-P*

S.N.: *4459*

Date: *AUG 18 1988*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN.
5 VOLTS/DIV
PRF ≈ 12 MHz

120/240 V
50-60 Hz