



PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-V-14V3-PS-P-CA-EN
 S.N.: 4775
 Date: JUL 30 83



Ⓐ 60 dB ATTEN
 100 VOLT/DIV
 5 KHz



Ⓑ AS Ⓐ BUT
 1 nSEC/DIV
 (RISE TIME)

- a) Output signal amplitude:
0 TO 100 VOLTS
- b) Pulse width:
3 TO 100 nSEC
- c) Rise time:
≤ 200 pSEC
- d) Fall time:
≤ 3 nSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
- g) Prime power:
OK

120/240V
 50-60 Hz
 [Signature]