


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APR-4B-PW-PS-PN*


S.N.: *4483*

Date: *1974 22 88*



- a) Output signal amplitude:  
*0 TO ± 400 VOLTS*
- b) Pulse width:  
*0.1 TO 100 USEC*
- c) Rise time:  
*≤ 20 NSEC*
- d) Fall time:  
*≤ 20 NSEC*
- e) PRF:  
*MIN PW: 0 TO 1.0 KHz*  
*MAX PW: 0 TO 50 Hz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*P<sub>OUT</sub> R<sub>L</sub> = 50 Ω*  
*PRF = 1.0 KHz*



*N<sub>OUT</sub> R<sub>L</sub> = 50 Ω*  
*PRF = 50 Hz*

*[Signature]*