


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-4B-PW-C-P*

S.N.: *4485*

Date: *DEC 9 '88*



a) Output signal amplitude:  
*0 TO +400 VOLTS*

b) Pulse width:  
*0.1 TO 100 µSEC*

c) Rise time:  
*≤ 20 NSEC*

d) Fall time:  
*≤ 20 NSEC*

e) PRF:  
*MAX PW: 0 TO 50 KHz*  
*PW < 5 µSEC: 0 TO 1 KHz*


f) Jitter, stability:

*OK*

g) Prime power:  
*120/240V*  
*50-60 KHz*

*[Signature]*

Ⓐ *R<sub>L</sub> = 50 Ω*  
*PRF ≈ 50 KHz*



Ⓑ *R<sub>L</sub> = 50 Ω*  
*PRF ≈ 1.0 KHz*