


PULSE GENERATOR
PERFORMANCE CHECK


Model: *APR-1-pw-CAN-05*

S.N.: *4489*

Date: *SEPT 13 88*



*P_{out} R_L = 50Ω
1.0 KHz PWF*



*N_{out} R_L = 50Ω
1.0 KHz PWF*

- a) Output signal amplitude:
0 TO ±200 VOLTS
- b) Pulse width:
0.1 TO 5.0 μSEC.
- c) Rise time:
≤ 10 nSEC
- d) Fall time:
≤ 10 nSEC
- e) PRF:
0 TO 1 KHz.
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
5.0 - 60 Hz.*