

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMT-5-1-07*

S.N.: *4492*

Date: *SEPT 22 88*



- a) Output signal amplitude: *0 to +10 Vpk*
- b) Pulse width: *≤ 130 psec*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:

4066 19720
10 Vpk/pk

120/240 V
50-60 Hz

[Signature]