



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-MV-2-C-P-M3 - J.M*
 S.N.: *4517*
 Date: *SEPT 27 88*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
A) 5 TO 50 NSR AND
B) 20 TO 500 NSR
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
A) 0 TO 1.0 MHz
B) 0 TO 200 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

A out. *40 db ATTEN*
10 VOLTS/DIV
100 MHz



B out *40 db ATTEN*
10 VOLTS/DIV
100 kHz

[Signature]