


PULSE GENERATOR
PERFORMANCE CHECK


Model: *APPP-2-C-P-PN-M*

S.N.: *4518*

Date: *OCT 21 88*



*A out 40 db ATTEN
10 VOLTS/DIV
1.0 MHz*



*B out 40 db ATTEN
10 VOLTS/DIV
1.0 MHz*

- a) Output signal amplitude:
0 TO ± 20 VOLTS
- b) Pulse width:
0.4 TO 100 nSEC
- c) Rise time:
A) 0.4 TO 50 nSEC B) 5 TO 100 nSEC
- d) Fall time:
≤ 200 pSEC
- e) PRF:
≤ 300 pSEC
- f) Jitter, stability:
OK
- g) Prime power:
0 TO 1.0 MHz

*120/240V
50-60 Hz*