


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P*
S.N.: *4541*
Date: *AUG 22 88*



- a) Output signal amplitude: *+25 TO +100. AMI*
- b) Pulse width: *≈ 25 NSEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 10 KHZ*
- f) Jitter, stability: *OK*
- g) Prime power:

≈ 60 AMPS/DIV +15 ± 1.0 VOLTS
500 HZ
200 mA (max)
[Signature]