


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P*

S.N.: *4542*

Date: *AUG 22 88*



- a) Output signal amplitude: *+25 TO +100 AMPS*
- b) Pulse width: *< 25 NSEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 100 KHZ*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 ± 1.0 VOLTS*

≈ 60 AMPS/DIV
500 Hz

+15 ± 1.0 VOLTS
200 mA MAX