


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P*
S.N.: *4543*
Date: *AUG 22 88*



- a) Output signal amplitude:
+ 25 TO +100 AMV.
- b) Pulse width:
≤ 25 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 KHz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 60 AMPS/DIV *+15 ± 1.0 VOLTS*
500 Hz. *200 mA (MAX)*

[Signature]