

PULSE GENERATOR
PERFORMANCE CHECK

Model:


NH-F

S.N.:

4545

Date:

OCT 25 1988



50 db ATTEN:

≈ 32 VOLTS/DIV

500 KHz (PRF)

a) Output signal amplitude:

0 TO +30 VOLTS

b) Pulse width:

≤ 100 NSEC.

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 100 MHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 VOLT

200 mA (MAX)