


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-2-PS-M-P*

S.N.: *4548*

Date: *Aug 19 1988*



- a) Output signal amplitude:
0 TO +350 VOLTS
- b) Pulse width:
5 TO 100 NSEC
- c) Rise time:
≤ 2 NSEC
- d) Fall time:
≤ 3 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
- g) Prime power:
OK

170 db ATTEN

~ 320 VOLTS/DIV

2 KHZ

120/240 V

50-60 Hz