


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PUH-5-P*
S.N.: *4551*
Date: *Jul 29 88.*



- a) Output signal amplitude:
0 TO +15 VOLT
- b) Pulse width:
≤ 500 PSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS / DIV
500 kHz (PRF)

+15 ± 1.0 VOLT
150 mA (MAX)

[Signature]