


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANNA-2-CP*

S.N.: *4552*

Date: *SEPT 28 '88*



40 db ATTEN
10 0045/DIN
1.0 AMPER PULS.

- a) Output signal amplitude: *0 TO +20 VOLTS*
- b) Pulse width: *20 TO 200 NSEC*
- c) Rise time: *(DUTY CYCLE MAX 20%)*
≤ 3 NSEC
- d) Fall time: *≤ 3 NSEC*
- e) PRF: *0 TO 10 MHz*
(20% MAX DUTY CYCLE)
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60 MHz

[Signature]