


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVR-4B-C-DN*

S.N.: *4554*

Date: *JAN 22 1996*



⊙ $P_{out} R_L = 50 \Omega$



⊙ $N_{out} R_L = 50 \Omega$

- a) Output signal amplitude:
0 TO $\pm 400 V$
- b) Pulse width:
 *$R_L \geq 50 \Omega$
0.1 TO 100 μs*
- c) Rise time: *(0.5% MAX DUTY CYCLE)*
 $\leq 20 ns$
- d) Fall time:
 $\leq 20 ns$
- e) PRF: *0 TO 1 KHz*
(0.5% MAX DUTY CYCLE)
- f) Jitter, stability: *OK*
- g) Prime power:
120/240 V
50-60 Hz

