


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-3-PS-PN-05-MITA*


S.N.: *4559*

Date: *NOV 2 88*




- a) Output signal amplitude: 0 TO ± 160 VOLTS
- b) Pulse width: TO $R_L > 100K$
 $1 \mu SEC$ TO DC
- c) Rise time: ≤ 70 NSEC
- d) Fall time: ≤ 10 NSEC
- e) PRF: 0 TO 20 KHz

Ⓐ $R_L > 100$ K OHM
TRIGGER AS PER
FIG 1



- f) Jitter, stability: OK
- g) Prime power: $120 / 240 \sqrt{30-60}$ Hz



Ⓑ AS Ⓐ BUT
 $2 \mu SEC / DIV$

Ⓒ AS A BUT SWITCH
IN DC POSITION.