


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AMR-3-PN-CA-KS1*

S.N.: *4569*

Date: *SEPT 21 88*



Ⓐ *N out (TOP) $R_L = 50 \Omega$
P out (BOT) $PRR = 500 \text{ Hz}$*



Ⓑ *AS Ⓐ BUT
 $PRF = 10 \text{ Hz}$
d PW NEAR MAX.*

- a) Output signal amplitude:
0 TO ± 200 VOLTS
- b) Pulse width: ($R_L = 50 \Omega$)
0.1 TO 2.50 μ SEC
- c) Rise time:
 $\leq 10 \text{ NSEC}$
- d) Fall time:
 $\leq 10 \text{ NSEC}$
- e) PRF:
0 TO 10 Hz
- f) Jitter, stability:
OK
- g) Prime power:
*+24 ± 1.0 VOLTS
1.0 AMP (MAX)*