


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVMR-1-P-CA-EW-ED-TRF

S.N.: 4571 - FL-M1-IM

Date: OCT 20 88



40 db ATTEN

5 VOLTS DIV

7.0 MHz

- a) Output signal amplitude: 0 TO +5 VOLTS
- b) Pulse width: 5 TO 50 nSEC
- c) Rise time: ≤ 100 pSEC
- d) Fall time: (OR ≤ 1 nSEC) ≤ 135 pSEC
- e) PRF: (OR ≤ 1 nSEC) 0 TO 10 MHz (20% MAX DUTY)
- f) Jitter, stability: CYCLE
- g) Prime power: OK
+24 \pm 1.0 VOLT
300 mA (MAX)

[Signature]