


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANMR-1-P-ET EW-ED - TRF*  
*- GCL-M1-JM*

S.N.: *4572*

Date: *OCT 20 88*



*40 dB ATTEN*  
*5 VOLTS/DIV*  
*7.0 MHz*

- a) Output signal amplitude:  
*0 TO +5 VOLTS.*
- b) Pulse width:  
*5 TO 50 NSEC*
- c) Rise time:  
*≤ 100 PSEC OR*
- d) Fall time: *≤ 1 NSEC*  
*≤ 135 PSEC OR*
- e) PRF: *≤ 1 NSEC*  
*0 TO 10 MHz*
- f) Jitter, stability: *(20% MAX DUTY CYCLE)*  
*DJK*
- g) Prime power:

*+24 ± 1.0 VOLTS*  
*300 MA (MAX)*

*[Signature]*