


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AMP-1-C-7-CA-EW*

S.N.: *4577*

Date: *NOV 16 88*



*A out 40 db ATTEN
10 VOLTS/DIV
1.0 MHz*



*B out 40 db ATTEN
10 VOLTS/DIV
1.0 MHz*

- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
0.2 TO 100. NSER
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*120/240 V
50 - 60 MHz*