


PULSE GENERATOR
PERFORMANCE CHECK


Model: *PVL-2A.-C-CMM1-P*

S.N.: *4583*

Date: *OCT 17 1988*



Ⓐ 60 db ATTEN
100 VOLT/DIV
PRF = 5.0 KHz



Ⓑ AS Ⓐ BUT
PRF = 20 KHz

- a) Output signal amplitude:
0 TO +170 V
- b) Pulse width:
5 TO 100 NSER
- c) Rise time:
6 2 NSER
- d) Fall time:
5 3 NSER
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/200 V
50-60 Hz