

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AVO - 5 - C - EA - EW - M - P - 05

S.N.:

4596

Date:

OCT 31 1988



60 db ATTEN

100 VOLTS/DIV

$R_L = 3 \Omega$

PRF = 2.0 KHz

a) Output signal amplitude:

0 TO +28 AMP

b) Pulse width:

6 TO 100 nSEC

c) Rise time:

≤ 3 nSEC

d) Fall time:

≤ 4 nSEC

e) PRF:

0 TO 5 KHz

f) Jitter, stability:

OK

g) Prime power:

120 / 240 V

50 - 60 Hz