


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-M-15-C-PN-0T*


S.N.: *4600*

Date: *JAN 16 '89*



- a) Output signal amplitude:
0 TO ±5 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 60 PSEZ
- d) Fall time:
≤ 130 PSEZ

*ⓐ P_{out} 40dB ATTEN
-1.5 VOLTS/DIV
1.0 MHz*



- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V

*ⓑ AS ⓐ BUT
N OUT*

*ⓐ OFFSET: 0 to ±5V
TO 50 Ω*

[Signature]