


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVPP-1-E-P-N-0T*

S.N.: *4601*

Date: *NOV 23 88*



- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
0.2 TO 100 NS

PW RANGE A *40db ATTEN*
10 VOLTS/DIV
1.0 MHz PRF

- c) Rise time:
± 100 PSEC
- d) Fall time:
± 135 PSEC



- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

PW RANGE B

40 db ATTEN
10 VOLTS/DIV
1.0 MHz PRF

120/240V
50-60 Hz