

PULSE GENERATOR
PERFORMANCE CHECK

Model:


PR-4A-PW-PS-P-CRI*

S.N.:

4606

Date:

NOV 25 88



$R_L = 50 \Omega$

PRF = 1 kHz

a) Output signal amplitude:

0 TO +400 V TO

b) Pulse width: *

50 ns

0.1 TO 3.0 μ SEC

c) Rise time:

≤ 20 NSEC

d) Fall time:

≤ 20 NSEC

e) PRF:

0 TO 1 kHz

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50-60 Hz

* CRI OPTION

SET PW RANGE AT

0.1 TO 3.0 μ SEC

RATHER THAN 5 NSEC

0.1 TO 5.0 μ SEC.