


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVR-A-1-PW-C-0A-EW-DN*

S.N.: *4610*

Date: *DEC 2 88*



*P<sub>out</sub> PRF = 5 KHz*



*Nowt PRF = 5 KHz*

- a) Output signal amplitude:  
*0 to ± 200 V<sub>0.2A</sub>*
- b) Pulse width:  
*50 nsec to 1.0 μsec\**
- c) Rise time:  
*± 10 nsec*
- d) Fall time:  
*± 10 nsec*
- e) PRF:  
*0 to 100 KHz\**
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 V*  
*50 - 60 Hz*

*\* DUTY CYCLE*  
*0.5% (MAX)*