


PULSE GENERATOR
PERFORMANCE CHECK


Model: *MM-1-C-N-PN-CA-LW-ED-M.*

S.N.: *4611*

Date: *NOV 17 88*



Ⓐ *40 db ATTEN*
5 VOLTS/DIV
≈ 25 MHz



Ⓑ *40 db ATTEN*
1.5 VOLTS/DIV
≈ 12 MHz

- a) Output signal amplitude:
0 TO 7.5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]